

Certificate of Analysis

Material: 86793320 **Benzethonium chlorure 0.05 mol/l (0.05N)**

Batch : 2500782 **Expires end of** 04/03/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,0499 --> 0,0501 mol/l	0,04988 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

10/03/2025

Signé par:

O. BRINGER , Responsable Qualité