

Certificate of Analysis

Material: 90196290 **Acide sulfurique 4,5 mol/l (9 N) en solution aqueuse**

Batch : 2500895 **Expires end of** 13/03/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	4,491-->4,509 mol/l	4,506 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

17/03/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité