

Certificate of Analysis

Material: MC3023321000 **Sodium thiosulfate 0.2 mol/l (0,2 N) en solution aqueuse**

Batch : 2500910 **Expires end of** 14/03/2026

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	0,1996 -->0,2004 mol/l	0,2003 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

17/03/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité