

Certificate of Analysis

Material: MC3116491000 **Cérium (IV) sulfate 0,05 mol/l (0,05 N)**

Batch : 2501150 **Expires end of** 02/04/2026

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	0,0499 --> 0,0501 mol/l	0,0499 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

09/04/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité