

# Certificate of Analysis

---

**Material:** MC3061195000      **Sodium hydroxyde 3 mol/l (3 N) en solution aqueuse**

**Batch :** 2501216      **Expires end of** 08/04/2026

---

<b>CHARACTERISTICS</b>	<b>SPECIFICATIONS</b>	<b>MESURED VALUES</b>
Titer (20°C; real value 0.2 % accuracy)	2,994 --> 3,006 mol/l	3,006 Mol/l

---

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**

**15/04/2025**

**Signé par:**



O. BRINGER , Responsable Qualité