

Certificate of Analysis

Material: 85103360 **EDTA sel disodique 0,2 mol/l (0,4 N) en solution aqueuse**

Batch : 2501253 **Expires end of** 09/04/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,1996 --> 0,2004 mol/l	0,2001 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

15/04/2025

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité