

# Certificate of Analysis

---

**Material:** MC3023225000      **Potassium permanganate 0.02 mol/l (0.1 N)  
en solution aqueuse**

**Batch :** 2501271      **Expires end of** 11/04/2026

<b>CHARACTERISTICS</b>	<b>SPECIFICATIONS</b>	<b>MESURED VALUES</b>
Titer (20°C; real value 0.2 % accuracy)	0,01996 --> 0,02004 mol/l	0,02000 Mol/l

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**

**15/04/2025**

**Signé par:**

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité