

Certificate of Analysis

Material: MC3116481000 **Sodium chlorure 0.01 mol/l (0.01 N) en solution aqueuse**

Batch : 2501296 **Expires end of** 14/04/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,00995-->0,01005 mol/l	0,00997 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

18/04/2025

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité