

Certificate of Analysis

Material: 87428290 **Sodium hydroxyde 9 mol/l (9 N) en solution aqueuse**

Batch : 2501298 **Expires end of** 14/04/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	8,99 --> 9,01 mol/l	9,00 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

18/04/2025

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité