

Certificate of Analysis

Material: MC3093820100 **Potassium permanganate 0.0025 mol/l
(0.0125 N) en solution aqueuse**

Batch : 2501491 **Expires end of** 02/05/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,00248-->0,00252 mol/l	0,002496 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

16/05/2025

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", with a long horizontal stroke extending to the right.

O. BRINGER , Responsable Qualité