

Certificate of Analysis

Material: 85495.9020 **Potassium hydroxyde 0,23 mol/l (0,23 N) en solution aqueuse**

Batch : 2501533 **Expires end of** 06/05/2026

| <i>CHARACTERISTICS</i> | <i>SPECIFICATIONS</i> | <i>MESURED VALUES</i> |
|---|-------------------------|-----------------------|
| Titer (20°C; real value 0.2 % accuracy) | 0,2296 --> 0,2304 mol/l | 0,2298 Mol/l |

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

16/05/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité