

# Certificate of Analysis

---

**Material:** MC3033631000      **Argent nitrate 0.2 mol/l (0.2 N) en solution aqueuse**

**Batch :** 2502027      **Expires end of** 16/06/2026

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	0,1996 --> 0,2004 mol/l	0,2002 Mol/l

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**

**25/06/2025**

**Signé par:**

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité