

Certificate of Analysis

Material: MC3075826010 **Acide nitrique 1,44 mol/l (1,44 N) en solution aqueuse**

Batch : 2502106 **Expires end of** 23/06/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	1,437-->1,443 mol/l	1,441 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

30/06/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité