

Certificate of Analysis

Material: MC3112672000 **Acide sulfurique 0,005 mol/l (0,01 N) en solution aqueuse**

Batch : 2502148 **Expires end of** 25/06/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,00499 --> 0,00501 mol/l	0,00500 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

30/06/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité