

Certificate of Analysis

Material: MC3129376010 **Acide sulfurique 0,13 mol/l (0,26 N) en solution aqueuse**

Batch : 2502168 **Expires end of** 26/06/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,1298-->0,1302 mol/l	0,1300 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

30/06/2025

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité