

Certificate of Analysis

Material: MC3023262500 **Argent nitrate 0.1 mol/l (0.1 N) en solution aqueuse**

Batch : 2502614 **Expires end of** 01/08/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,0998-->0,1002 mol/l	0,1001 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

07/08/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité