

Certificate of Analysis

Material: MC3052471000 **Acide chlorhydrique 2,5 mol/l (2,5 N)**

Batch : 2502856 **Expires end of** 29/08/2026

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	2,495-->2,505 mol/l	2,504 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

02/09/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité