

# Certificate of Analysis

---

**Material:** 89470290      **Acide chlorhydrique 3 mol/l (3N) +/- 0,2 %**

**Batch :** 2502951      **Expires end of** 05/09/2026

<b>CHARACTERISTICS</b>	<b>SPECIFICATIONS</b>	<b>MESURED VALUES</b>
Titer (20°C; real value 0.2 % accuracy)	2,994 --> 3,006 mol/l	3,004 mol/l

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**

**08/09/2025**

**Signé par:**



O. BRINGER , Responsable Qualité