

Certificate of Analysis

Material: MC2403101000 **Acide sulfurique 0,025 mol/l (0,05 N) en solution aqueuse**

Batch : 2502972 **Expires end of** 08/09/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,02495 --> 0,02505 mol/l	0,02500 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

15/09/2025

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité