

Certificate of Analysis

Material: MC3023321000 **Sodium thiosulfate 0.2 mol/l (0,2 N) en solution aqueuse**

Batch : 2502998 **Expires end of** 10/09/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,1996 -->0,2004 mol/l	0,1998 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

15/09/2025

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité