

Certificate of Analysis

Material: MC3129379020 **Acide sulfurique 0,13 mol/l (0,26 N) en solution aqueuse**

Batch : 2503007 **Expires end of** 10/09/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,1298-->0,1302 mol/l	0,1302 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

15/09/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité