

Certificate of Analysis

Material: MC3023250500 **Iode 0.05 mol/l (0.1 N) en solution aqueuse**

Batch : 2503081 **Expires end of** 17/09/2026

<i>CHARACTERISTICS</i>	<i>SPECIFICATIONS</i>	<i>MESURED VALUES</i>
Titer (20°C; real value 0.2 % accuracy)	0,0499 --> 0,0501 mol/l	0,0500 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

22/09/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité