

Certificate of Analysis

Material: 87628260 **Acide chlorhydrique 0.1 mol/l (0.1 N)**

Batch : 2503091 **Expires end of** 17/09/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,0998 --> 0,1002 mol/l	0,0998 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

22/09/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité