

Certificate of Analysis

Material: 90041290 **Potassium permanganate 0.004 mol/l (0.02 N) en solution aqueuse**
Batch : 2503432 **Expires end of** 16/10/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,003992--> 0,004008 mol/l	0,004004 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

20/10/2025

Signé par:



O. BRINGER , Responsable Qualité