

Certificate of Analysis

Material: 87128290 **Potassium hydroxyde 0,01 mol/l (0,01 N)**
dans le méthanol

Batch : 2503565 **Expires end of** 29/10/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,00995 -->0,01005 mol/l	0,01001 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

03/11/2025

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", with a long horizontal stroke extending to the right.

O. BRINGER , Responsable Qualité