

Certificate of Analysis

Material: MC3063731000 **Cérium (IV) sulfate 0,01 mol/l (0,01 N)**

Batch : 2503668 **Expires end of** 06/11/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Conforme au mode opératoire: 4,043 g cerium Sulfate Ref Prolabo 22604 10 ml acide sulfurique Ref Merck 100731 QSP 1000 ml eau désionisée Soit un titre = 0,01000 mol/l +/- 0,5 %	0,00996 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

17/11/2025

Signé par:



O. BRINGER , Responsable Qualité