

Certificate of Analysis

Material: MC3067615000 **Acide sulfurique 0,5 mol/l (1 N) en solution aqueuse**

Batch : 2503735 **Expires end of** 13/11/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,499-->0,501 mol/l	0,500 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

17/11/2025

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", with a long horizontal stroke extending to the right.

O. BRINGER , Responsable Qualité