

Certificate of Analysis

Material: 90482410 **Acide sulfurique 3,5 mol/l (7 N) en solution aqueuse**
Batch : 2503891 **Expires end of** 28/11/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	3,493 -->3,507 mol/l	3,499 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

01/12/2025

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité