

## Certificate of Analysis

---

**Material:** 85021.360 **Potassium permanganate 0.002 mol/l (0.01 N) en solution aqueuse**

**Batch :** 2503910 **Expires end of** 01/12/2026

| CHARACTERISTICS                         | SPECIFICATIONS            | MESURED VALUES |
|-----------------------------------------|---------------------------|----------------|
| Titer (20°C; real value 0.2 % accuracy) | 0.001996 - 0,002004 mol/l | 0,002000 mol/l |

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**

**09/12/2025**

**Signé par:**

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité