

Certificate of Analysis

Material: 90292410 **TRIS 0,025 N en solution aqueuse**

Batch : 2504011 **Expires end of** 08/12/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	Conforme au mode opératoire: 3.0285 g TRIS(HYDROXYMETHYL) AMINOMETHANE Ref VWR 103156 QSP 1000 ml eau désionisée Titre : 0.02495 -> 0.02505 N	0,02497 N

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

15/12/2025

Signé par:

O. BRINGER , Responsable Qualité