

Certificate of Analysis

Material: MC3006565000 **Potassium iodure 0,1 mol/l (0,1 N) en solution aqueuse**

Batch : 2504091 **Expires end of** 16/12/2026

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	16,6g Potassium iodure réf VWR 26846 QSP 1000 ml eau désionisée 0,0998 --> 0,1002 mol/l	0,0998 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

08/01/2026

Signé par:



O. BRINGER , Responsable Qualité