

Certificate of Analysis

Material: MC3052875000 **Argent nitrate 0.01 mol/l (0.01 N) en solution aqueuse**

Batch : 2600049 **Expires end of** 08/01/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,00995 --> 0,01005 mol/l	0,01000 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

26/01/2026

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité