

# Certificate of Analysis

---

**Material:** 87627260 **Acide chlorhydrique 0.02 mol/l (0.02 N)**

**Batch :** 2600446 **Expires end of** 11/02/2027

<b>CHARACTERISTICS</b>	<b>SPECIFICATIONS</b>	<b>MESURED VALUES</b>
Titer (20°C; real value 0.2 % accuracy)	0,01996 --> 0,02004 mol/l	0,02001 Mol/l

We Certify that this batch conforms to the specifications listed above  
BDL : Below detected limit

**Date**

**16/02/2026**

**Signé par:**

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité