

Certificate of Analysis

Material: 90213320 **Iode 0.025 mol/l (0.05 N) en solution aqueuse**

Batch : 2600550 **Expires end of** 19/02/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,02488 - 0,02512 mol/l	0,02504 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

23/02/2026

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité