

Certificate of Analysis

Material: 90277410 **EDTA sel disodique 0,001 mol/l (0,002 N) en solution aqueuse**

Batch : 2600775 **Expires end of** 09/03/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.2 % accuracy)	0,000995-->0,001005 mol/l	0,001001 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

17/03/2026

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité