

# Certificate of Analysis

**Material Number:** 85102.290  
**Material description :** Silver nitrate 0,01 mol/l (0.01 N) in 2 Propanol  
**Grade :** AVS TITRINORM®  
**Lot Number :** 2600781  
**Use by\* :** 2027-Mar-10  
**Date of Quality Control :** 2026-Mar-10  
**Date of Manufacture :** 2026-Mar-10  
**CAS Number :**  
**Molecular weight:**  
**Molecular formula:**

Characteristics	Specifications	Measured values
Titer (20°C; real value 0.2 % accuracy)	0.00998 -> 0.01002 mol/l	0,01001 mol/l

## Signature

We certify that this batch conforms to the specifications listed above. This document has been produced electronically and is valid without a signature.

Signed on behalf of VWR      2026-Mar-10  
O Bringer  
Quality Control Manager

## Other information

\*The use by date is the date to which VWR guarantees the above specification will be met for the product in an unopened container stored under recommended conditions.