

Certificate of Analysis

Material: MC3101761000 **Acide chlorhydrique 0.01 mol/l (0.01 N)**
Batch : 2601464 **dans l'isopropanol**
Expires end of 06/05/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,00995 --> 0,01005 mol/l	0,00999 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

19/05/2026

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité