

Certificate of Analysis

Material: 87128290 **Potassium hydroxyde 0,01 mol/l (0,01 N)**
Batch : 2601609 **dans le méthanol**
Expires end of 20/05/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,00995 -->0,01005 mol/l	0,01002 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

26/05/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité