

Certificate of Analysis

Material: MC3074211000 **Acide oxalique 0.5 mol/l (1 N) en solution aqueuse**

Batch : 2601650 **Expires end of** 26/05/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,499 --> 0,501 mol/l	0,501 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

01/06/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", with a long horizontal stroke extending to the right.

O. BRINGER , Responsable Qualité