

Certificate of Analysis

Material: 89579300 **Plomb nitrate solution 0.02 mol/l**

Batch : 2601682 **Expires end of** 28/05/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	Conforme au mode opératoire: 6,6242 g Plomb nitrate réf VWR 84847 QSP 1000 ml eau désionisée Titre=0,0200 mol/l +/-0,5 %	0,01996 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

01/06/2026

Signé par:



O. BRINGER , Responsable Qualité