

Certificate of Analysis

Material: MC3006541000 **Potassium permanganate 0.2 mol/l (1 N) en solution aqueuse**

Batch : 2601725 **Expires end of** 02/06/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,1996 -->0,2004 mol/l	0,2004 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

08/06/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité