

Certificate of Analysis

Material: MC3093371000 **Sodium thiosulfate 0.01 mol/l (0,01 N) en solution aqueuse**

Batch : 2601794 **Expires end of** 05/06/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,00998 --> 0,01002 mol/l	0,01002 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

08/06/2026

Signé par:

A handwritten signature in black ink, appearing to read 'O. Bringer', written over a horizontal line.

O. BRINGER , Responsable Qualité