

Certificate of Analysis

Référence : 89372.320

Solution d'acide citrique 3% à pH 8.0

Lot : 2601922

Date de validité

18/06/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Acide Citrique monohydraté Ph Eur Ref VWR 20276 lot 25L154121	30 g +/- 0.5%	conform
Eau HPLC Super Gradient Ref VWR 83650 lot 26A124017	QSP 1000 ml +/- 0.5%	conform
pH (Ajusté à 8.00 avec NaOH Ref 106498 lot MB2292998)	8.00 +/- 0.02	8.00
Filtration	0.22 µ	conform

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité