

Certificate of Analysis

Material: MC3052590100 **Acide sulfurique 2 mol/l (4 N) en solution aqueuse**

Batch : 2602037 **Expires end of** 29/06/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	1,996-->2,004 mol/l	2,001 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

06/07/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", with a long horizontal stroke extending to the right.

O. BRINGER , Responsable Qualité