

Certificate of Analysis

Material: MC303775.320 **Liqueur alcalimétrique HCl 0,02 N (N/50)**

Batch : 2602300 **Expires end of** 24/07/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,01996 --> 0,02004 mol/l	0,02000 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

27/07/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité