

Certificate of Analysis

Material: MC3099701000 **Cérium (IV) sulfate 0,1 mol/l (0,1 N)**

Batch : 2602380 **Expires end of** 31/07/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,0998 -->0,1002 mol/l	0,0999 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

04/08/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité