

Certificate of Analysis

Material: MC3067871000 **Acide nitrique 0,5 mol/l (0,5 N)**

Batch : 2602426 **Expires end of** 05/08/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,499 -->0,501 mol/l	0,499 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

10/08/2026

Signé par:

O. BRINGER , Responsable Qualité