

Certificate of Analysis

Material: MC3099729010 **Benzethonium chlorure 0.01 mol/l (0.01N)**

Batch : 2602498 **Expires end of** 11/08/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,00995 --> 0,01005 mol/l	0,01000 mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

17/08/2026

Signé par:

A handwritten signature in black ink, appearing to read "Bringer", with a long horizontal stroke extending to the right.

O. BRINGER , Responsable Qualité