

Certificate of Analysis

Material: 85021.360 **Potassium permanganate 0.002 mol/l (0.01 N) en solution aqueuse**

Batch : 2602533 **Expires end of** 17/08/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0.001996 - 0,002004 mol/l	0,001998 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

25/08/2026

Signé par:

A handwritten signature in black ink, appearing to read 'Bringer', with a long horizontal stroke extending to the right.

O. BRINGER , Responsable Qualité