

Certificate of Analysis

Material: MC3006490500 **Acide chlorhydrique 0.1 mol/l (0,1 N)**

Batch : 2602564 **Expires end of** 20/08/2027

CHARACTERISTICS	SPECIFICATIONS	MESURED VALUES
Titer (20°C; real value 0.3 % accuracy)	0,0998 --> 0,1002 mol/l	0,1001 Mol/l

We Certify that this batch conforms to the specifications listed above
BDL : Below detected limit

Date

25/08/2026

Signé par:

A handwritten signature in black ink, appearing to read "O. Bringer", written over a horizontal line.

O. BRINGER , Responsable Qualité